



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/087,205	02/28/2002	Michael L. Blomquist	14485.0147US01	2150
23552	7590	01/06/2009		
MERCHANT & GOULD PC P.O. BOX 2903 MINNEAPOLIS, MN 55402-0903			EXAMINER HANNE, SARA M	
			ART UNIT 2179	PAPER NUMBER
			MAIL DATE 01/06/2009	DELIVERY MODE PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

# Office Action Summary

**Application No.**

10/087,205

**Applicant(s)**

BLOMQUIST, MICHAEL L.

**Examiner**

SARA M. HANNE

**Art Unit**

2179

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 30 December 2008.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-16 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-16 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SE/US)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

### **DETAILED ACTION**

1. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on 12/30/08 has been entered. Claims 1-16 are currently pending.

#### ***Claim Rejections - 35 USC § 103***

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. Claims 1-4, 6-10 and 12-16 are rejected under 35 U.S.C. 103(a) as being unpatentable over Campbell et al., US Patent Application Publication 2005/0137530, hereinafter Campbell and further in view of Malave et al., US Patent Application Publication 2002/0193679, priority 9/29/1999, hereinafter Malave.

As in Independent Claims 1 and 15, Campbell teaches a pump and method for programming the pump comprising generating a user interface having a plurality of pages configured for display on a screen of a programmable pump, one of the pages being a home page (Page 1, par 6), a data port (cradle ref. 46), memory for storing a banner retrieved from the dataport (memory, ref. 16), retrieving a banner from memory of the programmable pump (Page 5, Par. 58), and displaying the retrieved banner in the

home page on the screen of the programmable pump (LCD, ref. 18, Fig. 22 status). While Campbell teaches the hardware and method of retrieving a banner and displaying in a home page on a screen, they fail to show the banner comprising user-defined patient-identifying content as recited in the claims. In the same field of the invention, Malave teaches a method for configuring a pump for delivering an agent to a patient similar to that of Campbell. In addition, Malave further teaches a user interface to allow a user to define a banner with patient-identifying parameter content to be displayed on the screen from memory (Fig. 13, patient name, Fig. 14 banner and corresponding text). It would have been obvious to one of ordinary skill in the art, having the teachings of Campbell and Malave before him at the time the invention was made, to modify the pump interface banner programming taught by Campbell to include the user-defined banner containing patient-identifying parameter content of Malave, in order to obtain a method for programming a pump to display a user defined banner containing the patient's name. One would have been motivated to make such a combination because a customizable banner of the user's choice for personal identification purposes would have been obtained, as taught by Malave.

As in Independent Claim 8, Campbell teaches a pump comprising memory for storing a banner (Fig. 1, ref. 16), a screen (Fig. 2, ref. 18), and a processor in data communication with the memory and screen (Fig. 1, ref. 14) programmed to retrieve the banner from memory (Page 5, Par. 58), and display it on the screen LCD, ref. 18, Fig. 22 status), wherein the banner identifies the pump as an insulin pump (Bolus data, Insulin type screen, U100U, Fig. 16 and 22). While Campbell teaches the hardware and

method of retrieving a banner and displaying in a home page on a screen, they fail to show the banner comprising user-defined patient-identifying content as recited in the claims. In the same field of the invention, Malave teaches a pump for delivering an agent to a patient similar to that of Campbell. In addition, Malave further teaches a user interface to allow a user to define a banner with patient-identifying content (Fig. 13, patient name, Fig. 14 banner and corresponding text). It would have been obvious to one of ordinary skill in the art, having the teachings of Campbell and Malave before him at the time the invention was made, to modify the pump interface banner programming, where the banner identifies the pump as an insulin pump taught by Campbell to include the user-defined banner containing patient-identifying content of Malave, in order to obtain a method for programming a pump to display a user defined banner containing patient-identifying content. One would have been motivated to make such a combination because a customizable banner of the user's choice for personal identification purposes would have been obtained, as taught by Malave.

As in Claim 2, Campbell teaches retrieving a user-defined banner identifying the programmable pump as an insulin pump (Bolus data, Insulin type screen, U100U, Fig. 16 and 22).

As in Claims 3 and 9, While Campbell teaches a programmable pump in data communication with a computer, the computer programmed to receive a banner through memory, they fail to show the banner containing information identifying the user of the programmable pump as recited in the claims. In the same field of the invention, Malave teaches a programmable pump similar to that of Campbell. In addition, Malave further

teaches a banner containing information identifying the user of the programmable pump (Fig 14, Patient Name). It would have been obvious to one of ordinary skill in the art, having the teachings of Campbell and Malave before him at the time the invention was made, to modify the programmable pump in data communication with a computer, the computer programmed to receive a banner from memory taught by Campbell to include the banner identifying the pump's user of Malave, in order to obtain a banner from memory displayed on the pump display identifying the pump's user. One would have been motivated to make such a combination because a personalized pump display would have been obtained, as taught by Malave.

As in Claims 4 and 10, Campbell teaches retrieving a banner containing medical information specific to the user of the programmable pump (Basal Review, Prime History).

As in Claims 6 and 12, Campbell teaches two or more of the pages generated by the user interface are home pages, and retrieving a banner from memory (ref. 16), includes retrieving two or more banners from memory, and displaying a banner one of the home pages and a banner on another of the home pages (Page 1, Par. 6 et seq.).

As in Claims 7 and 14, Campbell teaches downloading of the banner to the pump and loading the banner from memory onboard the programmable pump (Page 2, par 12-15).

As in Claim 13, Campbell teaches a data port, the processor being configured to receive a banner through the data port and store the banner in memory (See rejection of Claims 1 and 15 *supra*).

As in Independent Claim 16, Campbell teaches a data port (Fig. 1 ref. 46), memory (Fig. 1 ref. 16), the memory storing a list of selectively available banners (other users), a screen (Fig. 2 ref. 18), a processor in data communication with the data port (Fig. 1 ref. 14), the memory and the screen, the programmed to generate a user interface having a plurality of pages configured for display on the screen of the pump, at least one being a home page (Page 1, par 6), map one of the banners from the list of selectively available banners to the home page (Page 5, Par. 58), and display the banner in the home page (Fig. 22 status). While Campbell teaches the pump with hardware and programming to generate a home page which a banner is mapped to and displayed, they fail to show the banner comprising user-defined patient-identifying content as recited in the claims. In the same field of the invention, Malave teaches a method for configuring a pump for delivering an agent to a patient similar to that of Campbell. In addition, Malave further teaches a user interface to allow a user to define a banner with patient-identifying content (Fig. 13, patient name, Fig. 14 banner and corresponding text). It would have been obvious to one of ordinary skill in the art, having the teachings of Campbell and Malave before him at the time the invention was made, to modify the pump with hardware and programming to generate a home page which a banner is mapped to and displayed taught by Campbell to include the user-defined banner containing patient-identifying content of Malave, in order to obtain the pump with hardware and programming to generate a home page which a banner is mapped to and displayed, the banner containing patient-identifying content. One would have been motivated to make such a combination because a customizable banner of

the user's choice for personal identification purposes would have been obtained, as taught by Malave.

4. Claim 5 and 11 rejected under 35 U.S.C. 103(a) as being unpatentable over Campbell et al., US Patent Application Publication 2005/0137530, hereinafter Campbell, Malave et al., US Patent Application Publication 2002/0193679, priority 9/29/1999, hereinafter Malave and further in view of Estes et al., US Patent Application Publication 2003/0114836, hereinafter Estes.

Campbell and Malave teach a method for programming a pump to display a user defined banner containing non-pump parameter content (See claim 1 rejection *supra*). While Campbell and Malave teaches a programmable pump in data communication with a computer, the computer programmed to receive a banner through memory, the banner being user defined and containing patient-identifying content they fail to show retrieving a banner identifying the caregiver for the user of the programmable pump as recited in the claims. In the same field of the invention, Estes teaches a programmable pump similar to that of Campbell and Malave. In addition, Estes further teaches a banner containing information identifying the caregiver of the programmable pump (Fig 3A, Facility Info, Physician Name). It would have been obvious to one of ordinary skill in the art, having the teachings of Campbell and Malave and Estes before him at the time the invention was made, to modify the programmable pump in data communication with a computer, the computer programmed to receive a banner from memory, the banner being user defined and containing patient-identifying content taught by Campbell and Malave to include the banner identifying the pump's caregiver of Estes, in order to



obtain a banner from memory, the banner being user defined and containing patient-identifying content displayed on the pump display identifying the pump's caregiver. One would have been motivated to make such a combination because an authorized pump display would have been obtained, as taught by Estes.

***Response to Arguments***

Applicant's arguments filed 12/30/08 have been fully considered but they are not persuasive.

The examiner notes that the claims do not define a freely editable banner or entering any information "as desired", but merely recite that the banner displays user-defined non-pump parameter content. If the applicant is equating the non-pump parameter to the owner id of the pump, the examiner notes that Malave teaches the user choosing the owner id fields (Fig. 12) and displaying it in the banner (Fig. 14).

In response to the applicant's arguments regarding Campbell's teachings, the examiner agrees that Campbell does not teach patient-identifying information displayed in the banner of the pump, however this limitation is taught by Malave.

In response to the applicant's arguments regarding Malave, the examiner disagrees. Malave clearly teaches a banner with patient identifying data retrieved from memory as seen in the Figures and referenced in the claim rejections above. Furthermore the transmission of data from the user on a PC to a pump is disclosed by Campbell as cited above. The combination of Campbell and Malave would result in the claim limitations presented and is merely a software issue that is not new.

***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to SARA M. HANNE whose telephone number is (571)272-4135. The examiner can normally be reached on M-F 7:30am-4:00pm, off on alternating Fridays.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, WEILUN LO can be reached on (571) 272-4847. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Sara M Hanne/  
Examiner, Art Unit 2179